

Worldwide Headquarters • Esco Micro Pte. Ltd. • 21 Changi South Street 1 • Singapore, 486777
Phone +65 6542 0833 • Fax +65 6546 2913, +65 6542 5732, +65 6542 6920 • mail@escoglobal.com • www.escoglobal.com

ROC No. 19840016SW


Declaration of Conformation

In accordance to EN ISO/IEC 17050-1:2010

We, Esco Micro Pte. Ltd.

of 21 Changi South Street 1

Singapore, 486777 Tel: +65 6542 0833 Fax: +65 6542 6920



declare on our sole responsibility that the product:

Category : Class II Biological Safety Cabinets

Brand : Cytoculture Model : LS2-4A1

in accordance with the following directives:

• 2006/95/EEC : The Low Voltage Directive and its amending directives

• 2004/108/CE : The Electromagnetic Compatibility Directive and its amending

directives

has been designed to comply with the requirement of the following Harmonized Standard:

• Low Voltage : EN 61010-1:2010

• EMC : EN 61326-1:2006 Class B

• Design/ : EN 12469:2000 Class II Biological Safety Cabinet

Performance Criteria

More information may be obtained from Esco's authorized distributors located within the European Union. A list of these parties and their contact information is available on request from Esco.

XQ Lin Group CEO, Esco (Feb 14th, 2011)